

Operator: Validation Manager
Process: Validation
Session Start: 26-11-2013 12:52:20
Session Stop: 26-11-2013 12:52:20
Session Name: Sterilization Validation
Session Text:
Vessel: Sterilizer
Product: Pharma Medical
Time Zone: UTC offset 01:00:00



Lethality Overview

Input parameter

Start Date : 03-04-2008
Ref. Temperature: 121,10 °C
Min. Temperature : 91,10 °C
Z-Value: 10 °C
D-Value:

Sensor lethality overview

Name	Description	Start Time	End Time	Max. Value
Lethality - 1		13:10:22	13:21:22	222,83
Lethality - 2		13:10:22	13:21:22	222,86
Lethality - 3		13:10:22	13:21:22	221,98
Lethality - 4		13:10:22	13:21:22	222,18
Lethality - 5		13:10:22	13:21:22	221,79
Lethality - 6		13:10:22	13:21:22	222,85
Lethality - Afløb		13:10:22	13:21:22	221,25